English

Wrist and Forearm Set

INSTRUCTIONS FOR USE (IFU) FOR THE UNITED STATES OF AMERICA

℞: For use by physicians only. Federal Law (USA) restricts this device to sale by or on order of a physician.

This package insert is designed to provide Instructions for Use of the Wrist and Forearm Set; it is not a reference to surgical techniques. **Failure to follow instructions may lead to patient injury.**

Description

The Skeletal Dynamics Wrist and Forearm Set consists of the following systems:

1. GEMINUS® Volar Distal Radius Plating System
2. PROTEAN® Fragment Plate System
3. Dorsal Spanning Plate System
4. Forearm Plating System

All components of the system are provided non-sterile to be sterilized in the user facility.

1. GEMINUS® Volar Distal Radius Plating System

The Skeletal Dynamics GEMINUS® Volar Distal Radius Plating System contains bone plates for the repair of distal volar radial fractures. Included in the set are titanium bone screws, fixation pegs, fragment plates, and specialized instrumentation. Also included are Hook Plate Extensions to buttress a volar marginal fragment, a Buttress Button Extension to buttress a dorsal marginal fragment, and cannulated cobalt chrome polyaxial locking screws for trajectories different than those of the fixed angled bone plates.

The GEMINUS® Volar Distal Radius Plates are available in various sizes and are made of medical-grade titanium alloy. Cortical screws affix the plate to the diaphysis, and fixed-angle pegs are used for distal bone fragments. The system is provided non-sterile and sterilized in the user facility.

The GEMINUS® Volar Distal Radius Plating System is comprised of:

- CP Titanium and Titanium alloy buttress button assembly
- Titanium alloy plates, washers, and screws
- CoCr Cannulated Polyaxial Locking Screw (PLS)
- Stainless steel K-Wires (for provisional fixation; not for implantation)
- System specific instrumentation

Indications for Use of GEMINUS Volar Distal Radius Plating System

The GEMINUS® Volar Distal Radius Plating System is intended for the fixation of fractures and osteotomies of the distal radius.

2. PROTEAN® Fragment Plate System

The PROTEAN® Fragment Plate System is a Set of titanium alloy bone plates designed for stabilization and repair of small bone fragments. Included in the Set are titanium alloy bone screws and pegs, cobalt chrome cannulated polyaxial screws, K-Wires, and specialized instrumentation.

The PROTEAN® Fragment Plates are available in various configurations. The system is provided non-sterile and is sterilized in the user facility.

IFU-02499-00 V.1 – February 2026
© 2026 Skeletal Dynamics, Inc

The PROTEAN® Fragment Plate System is comprised of:

- Titanium alloy plates and screws
- Cobalt chrome polyaxial screws
- Stainless steel K-wires (for provisional fixation; not for implantation)
- System specific instrumentation.

Indications for Use PROTEAN® Fragment Plate System

The Skeletal Dynamics PROTEAN® Fragment Plate System is intended for stabilization and fixation of small bone fragments in fresh fractures, revision procedures, joint fusion and reconstructions of small bones of the hand, foot, wrist, ankle, humerus, scapula, finger, toe, and pelvis, particularly in osteopenic bone.

3. Dorsal Spanning Plate System

The Skeletal Dynamics Dorsal Spanning Plate System contains bone plates for the repair of distal radius fractures and wrist arthrodesis. Included in the Set are titanium bone screws, plates, and specialized instrumentation.

The plates are available in two sizes and made of medical grade titanium alloy. Cortical screws affix the plate to the diaphysis of the bone. The system is provided non-sterile and is sterilized in the user facility.

The Dorsal Spanning Plate System is comprised of:

- Titanium alloy plates and screws
- System Specific instrumentation

Indications for Use Dorsal Spanning Plate System

The Skeletal Dynamics Dorsal Spanning Plate System is intended for the fixation of fractures involving the distal radius and for wrist arthrodesis.

4. Forearm Plating System

The Forearm Plating System contains bone plates for the repair of midshaft radius and ulna fractures. Included in the Set are titanium alloy bone screws, K-Wires, and specialized instrumentation.

The Forearm Plates are available in various sizes and are made of medical grade titanium alloy. The system is provided non-sterile and is sterilized at the user facility.

The Forearm Plating System is comprised of:

- Titanium alloy plates and screws
- Stainless steel K-Wires (for provisional fixation; not for implantation
- System specific instrumentation

Indications for Use Forearm Plating System

The Skeletal Dynamics Forearm Plating System is indicated for the treatment of fractures, fusions, and osteotomies of the radius and ulna.

Contraindications

Prior to using the Wrist and Forearm Set, ensure that none of the following patient conditions are present: active or latent infection, sepsis, insufficient quantity or quality of bone and/or soft tissue, material sensitivity, or patients who are unwilling or incapable of following postoperative care instructions.

IFU-02499-00 V.1 – February 2026
© 2026 Skeletal Dynamics, Inc

⚠ General Warnings and Precautions

- The information in this document should be shared with the patient.
- The patient should be informed about the importance of following the post-operative rehabilitation to fully understand the limitations in activities of daily living.The patient must be warned that failure to follow postoperative care instructions may cause the implant or treatment to fail.
- The patient must be cautioned about the use, limitations, and potential adverse effects of this device including the possibility of delayed union, non-union, device or treatment failure as a result loosening, stress, excessive activity, or weight bearing or load bearing, and the possibility of nerve or soft tissue damage related to either surgical trauma or the presence of the device.
- Potential construct failures such as implant breakage, loosening, fixation failure, instability, delayed soft tissue healing, soft tissue irritation, delayed union, or non-union may occur as a result of noncompliance to post-operative rehabilitation, excessive activities, or construct overloading.
- For safe effective use of the implant, the surgeon must be thoroughly familiar with the surgical technique implant, and associated instruments.
- The device is not designed to withstand the stress excessive of weight bearing, load bearing, or physical activity. Improper application of the device during implantation may also increase the possibility of loosening, or migration.
- DO NOT reuse any of the Set implantable components. Reuse may compromise the structural integrity of the construct.
- DO NOT reuse instruments that are laser marked with the words “do not reuse" or corresponding symbol .
- Protect the Set’s implantable components against scratching or nicking. Such stress concentration can lead to implant failure.
- Before using the Set, inspect all implants and instruments for wear, disfiguration and physical damage. If evidence of wear, disfiguration or physical damage is found, DO NOT use and contact your local Skeletal Dynamics representative or the Skeletal Dynamics Customer Care Department.
- DO NOT permanently implant the Skeletal Dynamics K-Wires; they are only intended to be used during provisional fixation.
- DO NOT permanently implant the pre-loaded Drill Guides or Aiming Guides; they are intended to be removed prior to screw insertion.
- DO NOT use peg/screw lengths that will excessively protrude through the far cortex as it may result in soft tissue irritation.
- If the screws are not attached and/or fully tightened, a non-union, delayed union or construct failure may occur.
- DO NOT mix system specific instrumentation from different systems or manufacturers for metallurgical, biomechanical and functional reasons.
- Dispose of contaminated implants and instruments per established facility guidelines and protocols.
- Caution should be taken for interference to pacemakers during electrocautery or by uncertified drill power sources.
- Seek medical help immediately if implant malfunctions.
- The benefits from implant surgery may not meet the patient’s expectations or may deteriorate over time, requiring revision surgery to replace the implant or to carry out alternative procedures.
- Care should be taken that no implants are placed interfering with the articular surface.

© 2026 Skeletal Dynamics, Inc

- To maintain traceability of the implantable components, record each of the respective components Lot numbers in the patient records post implantation.

⚠ GEMINUS® Volar Plating System Warnings and Precautions

- GEMINUS® Drill Blocks are only compatible with GEMINUS® Volar Plates containing a Gold PDG in the shaft.
- Assure Peg Driver tip does not show any signs of wear or distress such as rounded square edges, excessive depth marks from peg recess insertion, or deformed twisted tip. If such evidence is found for Peg Driver, DO NOT USE and contact your local Skeletal Dynamics representative or the Skeletal Dynamics Customer Care Department for replacement.
- Use only one 2.5mm PLS in each head of the GEMINUS® Volar Plate.
- The maximum angulation of the PLS should not exceed 10° from the trajectory of the respective hole.
- Use only one 2.7mm Peg (High Compression or Fully Threaded, Non-Locking) in each head of the GEMINUS® Volar Plate.
- DO NOT use the PLS in the most distal hole(s) on the lunate head of the GEMINUS® Volar Plate.
- DO NOT open the volar capsule as it may devascularize fracture fragments and destabilize the volar wrist ligaments.
- The Non-locking Threaded Pegs are NOT intended to provide subchondral support. Their use should be limited to capture remote bone fragments where partially or fully threaded pegs cannot be used.
- The use of power tools for the installation of the screws and pegs is not recommended and may lead to cross treading and damage to the screws and/or plates.

⚠ PROTEAN® Fragment Plate System Warnings and Precautions

- The maximum angulation of the PLS should not exceed 10° from the trajectory of the respective hole.
- The Non-locking Threaded Pegs are NOT intended to provide subchondral support. Their use should be limited to capture remote bone fragments where partially or fully threaded pegs cannot be used.
- DO NOT open the volar capsule as it may devascularize fracture fragments and destabilize the volar wrist ligaments.

⚠ Dorsal Spanning Plate System Warnings and Precautions

- If used for fracture fixation, the plate must be removed after fracture healing.

Potential Adverse Effects

The following are potential risks that have been associated with wrist surgery: infection, nonunion, persistent pain, stiffness of the fingers, loosening or migration of the implants resulting in misalignment.

MR Safety Information

The devices included in the Wrist and Forearm Set have not been evaluated for safety and compatibility in the MR environment. They have not been tested for heating, migration, or image artifact in the MR environment. The safety of the Wrist and Forearm Set in the MR environment is unknown. Scanning a patient who has this device implanted may result in patient injury.

© 2026 Skeletal Dynamics, Inc

Directions for Use

The Wrist and Forearm Set should only be used by surgeons who have experience with this system being used. Each surgeon must evaluate the appropriateness for the use of the Wrist and Forearm Set based on their clinical experiences.

Please refer to the Surgical Technique Guide(s) (MKT-00005-00 & MKT-00253-00) to review the surgical approach as described by Jorge L. Orbay, M.D. of the Miami Hand and Upper Extremity Institute located in Miami, Florida, USA.

Cleaning, Sterilization, and Inspection

For instructions on cleaning, disinfection, sterilization and inspection of the Wrist & Forearm Set products please refer to the cleaning and sterilization instructions for use (IFU-04056-00). Refer to the Surgical Technique Guide(s)(MKT-00005-00 & MKT-00253-00) for proper kitted tray arrangement.

Resources

The latest version of the Instructions for Use may be requested in a physical format by email (orders@skeletaldynamics.com) or by phone (+1-877-753-5396). The physical copy will be provided within 7 calendar days of receiving a request from the user or at the time of delivery of the device, if so, requested at the time of order.

For the most current instructions for use visit www.skeletaldynamics.com/resources. Instructions for Use should always be reviewed before using or implanting a device.

Disclaimer of Warranty and Limited Remedies

Skeletal Dynamics, Inc makes no express or implied warranty, including any implied warranty of merchantability or fitness for a particular purpose, on the product(s) described in this publication. Skeletal Dynamics, Inc shall not be liable under any circumstances for any direct, incidental or consequential damages other than as expressly provided by specific law. No person has authority to bind Skeletal Dynamics, Inc to any representation or warranty except as specifically set forth in this publication. Descriptions or specifications provided by Skeletal Dynamics, Inc in any publication are only included to generally describe the product when manufactured and do not constitute any express warranties.



Skeletal Dynamics, Inc
7300 N. Kendall Dr. / Suite 800
Miami, FL 33156
United Stated of America
1-877-753-5396



IFU-02499-00 V.1 – February 2026
© 2026 Skeletal Dynamics, Inc

SYMBOLS GLOSSARY

Symbols that follow BS EN ISO 15223-1 / Medical Devices-Symbols to be used with medical device labels, labelling and information to be supplied.

Symbol	Symbol Reference Number and Title	Description
	5.2.7 Non-sterile	Indicates a medical device that has not been subjected to a sterilization process
	5.2.6 Do not resterilize	Indicates a medical device that is not to be resterilized
	5.1.6 Catalogue Number	Indicates the manufacturer's catalogue number so the medical device can be identified
	5.4.3 Consult Instructions for Use	Indicates the need for the user to consult the instructions for use
	5.1.5 Batch Code	Indicates the manufacturer's batch code or lot can be identified
	5.4.2 Do Not Re-use	Indicates a medical device that is intended for one single use only
	5.1.1 Manufacturer	Indicates the medical device manufacturer
	5.2.8 Do Not Use If Package is Damaged	Indicates a medical device that should not be used if the package has been damaged or opened
	5.1.4 Use-by-date	Indicates the date after which the medical device is not to be used
	5.7.10 Unique Device Identifier	Indicates a carrier that contains unique device identifier information
	5.7.7 Medical device	Indicated the item is a medical device
	5.1.3 Date of Manufacture	Indicates the date when the medical device was manufactured

© 2026 Skeletal Dynamics, Inc

Wrist and Forearm Set

Inventory Control Sheet

GEMINUS Volar Distal Radius Plates	
GEMINUS Plate, Narrow, 3 Hole, Right GMN-RTN-3HL (01)00841506101620 (01) 00841506101620	GEMINUS Plate, Narrow, 3 Hole, Left GMN-LTN-3HL (01)00841506101569 (01) 00841506101569
GEMINUS Plate, Narrow, 4 Hole, Right GMN-RTN-4HL (01)00841506101637 (01) 00841506101637	GEMINUS Plate, Narrow, 4 Hole, Left GMN-LTN-4HL (01)00841506101576 (01) 00841506101576
GEMINUS Plate, Standard, 3 Hole, Right GMN-RTS-3HL (01)00841506101644 (01) 00841506101644	GEMINUS Plate, Standard, 3 Hole, Left GMN-LTS-3HL (01)00841506101583 (01) 00841506101583
GEMINUS Plate, Standard, 4 Hole, Right GMN-RTS-4HL (01)00841506101651 (01) 00841506101651	GEMINUS Plate, Standard, 4 Hole, Left GMN-LTS-4HL (01)00841506101590 (01) 00841506101590
GEMINUS Plate, Standard, 7 Hole, Right GMN-RTS-7HL (01)00841506101668 (01) 00841506101668	GEMINUS Plate, Standard, 7 Hole, Left GMN-LTS-7HL (01)00841506101606 (01) 00841506101606
GEMINUS Plate, Wide, 4 Hole, Right GMN-RTW-4HL (01)00841506101675 (01) 00841506101675	GEMINUS Plate, Wide, 4 Hole, Left GMN-LTW-4HL (01)00841506101613 (01) 00841506101613
GEMINUS Plate, Narrow, 120mm, Right GMN-RTN-120 (01)00841506107455 (01) 00841506107455	GEMINUS Plate, Narrow, 120mm, Left GMN-LTN-120 (01)00841506107462 (01) 00841506107462
GEMINUS Plate, Narrow, 160mm, Right GMN-RTN-160 (01)00841506107479 (01) 00841506107479	GEMINUS Plate, Narrow, 160mm, Left GMN-LTN-160 (01)00841506107486 (01) 00841506107486
GEMINUS Plate, Narrow, 200mm, Right GMN-RTN-200 (01)00841506107493 (01) 00841506107493	GEMINUS Plate, Narrow, 200mm, Left GMN-LTN-200 (01)00841506107509 (01) 00841506107509
Midshaft Radius Plate	**Midshaft Ulna Plate**
FreeFix Midshaft Radius Plate, 81mm MRP-081 (01)00841506114576 (01) 00841506114576	FreeFix Midshaft Ulna Plate, 81mm MUP-081 (01)00841506115351 (01) 00841506115351
FreeFix Midshaft Radius Plate, 97mm MRP-097 (01)00841506112992 (01) 00841506112992	FreeFix Midshaft Ulna Plate, 97mm MUP-097 (01)00841506115368 (01) 00841506115368
FreeFix Midshaft Radius Plate, 112mm MRP-112 (01)00841506114583 (01) 00841506114583	FreeFix Midshaft Ulna Plate, 112mm MUP-112 (01)00841506115375 (01) 00841506115375
FreeFix Midshaft Radius Plate, 127mm MRP-127 (01)00841506113005 (01) 00841506113005	FreeFix Midshaft Ulna Plate, 127mm MUP-127 (01)00841506115382 (01) 00841506115382

IFU-02499-00 V.1 – February 2026
© 2026 Skeletal Dynamics, Inc

Wrist and Forearm Set
Inventory Control Sheet

FreeFix Midshaft Radius Plate, 142mm MRP-142 (01)00841506114590 (01)00841506114590	FreeFix Midshaft Ulna Plate, 142mm MUP-142 (01)00841506113036 (01)00841506113036
FreeFix Midshaft Radius Plate, 157mm MRP-157 (01)00841506113012 (01)00841506113012	FreeFix Midshaft Ulna Plate, 157mm MUP-157 (01)00841506115399 (01)00841506115399
FreeFix Midshaft Radius Plate, 173mm MRP-173 (01)00841506115733 (01)00841506115733	FreeFix Midshaft Ulna Plate, 173mm MUP-173 (01)00841506115405 (01)00841506115405
FreeFix Midshaft Radius Plate, 188mm MRP-188 (01)00841506115740 (01)00841506115740	FreeFix Midshaft Ulna Plate, 188mm MUP-188 (01)00841506115412 (01)00841506115412
PROTEAN Plates	
PROTEAN Fragment Plate, Radial Column Plate, Right PRT-RCP-RT (01)00841506109923 (01)00841506109923	PROTEAN Fragment Plate, Radial Column Plate, Left PRT-RCP-LT (01)00841506109930 (01)00841506109930
PROTEAN Fragment Plate, Central Column Plate, Right PRT-CCP-RT (01)00841506109947 (01)00841506109947	PROTEAN Fragment Plate, Central Column Plate, Left PRT-CCP-LT (01)00841506109954 (01)00841506109954
PROTEAN Fragment Plate, Distal Ulna PRT-FSP-DU (01)00841506102900 (01)00841506102900	PROTEAN Fragment Plate, Double Hockey Stick PRT-FSP-LR (01)00841506102917 (01)00841506102917
PROTEAN Fragment Plate, Y PRT-FSP-YS (01)00841506102931 (01)00841506102931	
Smooth Peg, Locking	
Smooth Peg, Locking, 2.0mm x 14mm, Ti SPLS-20140-TS (01)00841506102962 (01)00841506102962	Smooth Peg, Locking, 2.0mm x 21mm, Ti SPLS-20210-TS (01)00841506103020 (01)00841506103020
Smooth Peg, Locking, 2.0mm x 16mm, Ti SPLS-20160-TS (01)00841506102979 (01)00841506102979	Smooth Peg, Locking, 2.0mm x 22mm, Ti SPLS-20220-TS (01)00841506103037 (01)00841506103037
Smooth Peg, Locking, 2.0mm x 17mm, Ti SPLS-20170-TS (01)00841506102986 (01)00841506102986	Smooth Peg, Locking, 2.0mm x 23mm, Ti SPLS-20230-TS (01)00841506103044 (01)00841506103044
Smooth Peg, Locking, 2.0mm x 18mm, Ti SPLS-20180-TS (01)00841506102993 (01)00841506102993	Smooth Peg, Locking, 2.0mm x 24mm, Ti SPLS-20240-TS (01)00841506103051 (01)00841506103051
Smooth Peg, Locking, 2.0mm x 19mm, Ti SPLS-20190-TS (01)00841506103006 (01)00841506103006	Smooth Peg, Locking, 2.0mm x 26mm, Ti SPLS-20260-TS (01)00841506103068 (01)00841506103068

IFU-02499-00 V.1 – February 2026
© 2026 Skeletal Dynamics, Inc

Wrist and Forearm Set
Inventory Control Sheet

Smooth Peg, Locking, 2.0mm x 20mm, Ti SPLS-20200-TS (01)00841506103013 (01)00841506103013	Smooth Peg, Locking, 2.0mm x 28mm, Ti SPLS-20280-TS (01)00841506103075 (01)00841506103075
Threaded Peg, Locking	
Threaded Peg, Locking, 2.3mm x 10mm, Ti TPLS-23100-TS (01)00841506103358 (01)00841506103358	Threaded Peg, Locking, 2.3mm x 23mm, Ti TPLS-23230-TS (01)00841506103457 (01)00841506103457
Threaded Peg, Locking, 2.3mm x 12mm, Ti TPLS-23120-TS (01)00841506103365 (01)00841506103365	Threaded Peg, Locking, 2.3mm x 24mm, Ti TPLS-23240-TS (01)00841506103464 (01)00841506103464
Threaded Peg, Locking, 2.3mm x 14mm, Ti TPLS-23140-TS (01)00841506103372 (01)00841506103372	Threaded Peg, Locking, 2.3mm x 26mm, Ti TPLS-23260-TS (01)00841506103471 (01)00841506103471
Threaded Peg, Locking, 2.3mm x 16mm, Ti TPLS-23160-TS (01)00841506103389 (01)00841506103389	Threaded Peg, Locking, 2.3mm x 28mm, Ti TPLS-23280-TS (01)00841506103488 (01)00841506103488
Threaded Peg, Locking, 2.3mm x 17mm, Ti TPLS-23170-TS (01)00841506103396 (01)00841506103396	Threaded Peg, Locking, 2.3mm x 30mm, Ti TPLS-23300-TS (01)00841506103495 (01)00841506103495
Threaded Peg, Locking, 2.3mm x 18mm, Ti TPLS-23180-TS (01)00841506103402 (01)00841506103402	Threaded Peg, Locking, 2.3mm x 32mm, Ti TPLS-23320-TS (01)00841506103501 (01)00841506103501
Threaded Peg, Locking, 2.3mm x 19mm, Ti TPLS-23190-TS (01)00841506103419 (01)00841506103419	Threaded Peg, Locking, 2.3mm x 34mm, Ti TPLS-23340-TS (01)00841506103983 (01)00841506103983
Threaded Peg, Locking, 2.3mm x 20mm, Ti TPLS-23200-TS (01)00841506103426 (01)00841506103426	Threaded Peg, Locking, 2.3mm x 36mm, Ti TPLS-23360-TS (01)00841506103990 (01)00841506103990
Threaded Peg, Locking, 2.3mm x 21mm, Ti TPLS-23210-TS (01)00841506103433 (01)00841506103433	Threaded Peg, Locking, 2.3mm x 38mm, Ti TPLS-23380-TS (01)00841506104003 (01)00841506104003
Threaded Peg, Locking, 2.3mm x 22mm, Ti TPLS-23220-TS (01)00841506103440 (01)00841506103440	Threaded Peg, Locking, 2.3mm x 40mm, Ti TPLS-23400-TS (01)00841506104010 (01)00841506104010
High Compression Locking Peg	
High Compression Locking Peg, 2.7mm x 10mm, Ti HCLP-27100-TS (01)00841506101682 (01)00841506101682	High Compression Locking Peg, 2.7mm x 21mm, Ti HCLP-27210-TS (01)00841506101750 (01)00841506101750
High Compression Locking Peg, 2.7mm x 12mm, Ti HCLP-27120-TS (01)00841506101699 (01)00841506101699	High Compression Locking Peg, 2.7mm x 22mm, Ti HCLP-27220-TS (01)00841506101767 (01)00841506101767

© 2026 Skeletal Dynamics, Inc

Wrist and Forearm Set
Inventory Control Sheet

High Compression Locking Peg, 2.7mm x 14mm, Ti HCLP-27140-TS (01)00841506101705 (01)00841506101705	High Compression Locking Peg, 2.7mm x 23mm, Ti HCLP-27230-TS (01)00841506101774 (01)00841506101774
High Compression Locking Peg, 2.7mm x 16mm, Ti HCLP-27160-TS (01)00841506101712 (01)00841506101712	High Compression Locking Peg, 2.7mm x 24mm, Ti HCLP-27240-TS (01)00841506101781 (01)00841506101781
High Compression Locking Peg, 2.7mm x 18mm, Ti HCLP-27180-TS (01)00841506101729 (01)00841506101729	High Compression Locking Peg, 2.7mm x 26mm, Ti HCLP-27260-TS (01)00841506101798 (01)00841506101798
High Compression Locking Peg, 2.7mm x 19mm, Ti HCLP-27190-TS (01)00841506101736 (01)00841506101736	High Compression Locking Peg, 2.7mm x 28mm, Ti HCLP-27280-TS (01)00841506101804 (01)00841506101804
High Compression Locking Peg, 2.7mm x 20mm, Ti HCLP-27200-TS (01)00841506101743 (01)00841506101743	High Compression Locking Peg, 2.7mm x 30mm, Ti HCLP-27300-TS (01)00841506101811 (01)00841506101811
Threaded Peg, Non-Locking	
Threaded Peg, Non-Locking, 2.7mm x 10mm, Ti TPNL-27100-TS (01)00841506103518 (01)00841506103518	Threaded Peg, Non-Locking, 2.7mm x 26mm, Ti TPNL-27260-TS (01)00841506103594 (01)00841506103594
Threaded Peg, Non-Locking, 2.7mm x 12mm, Ti TPNL-27120-TS (01)00841506103525 (01)00841506103525	Threaded Peg, Non-Locking, 2.7mm x 28mm, Ti TPNL-27280-TS (01)00841506103600 (01)00841506103600
Threaded Peg, Non-Locking, 2.7mm x 14mm, Ti TPNL-27140-TS (01)00841506103532 (01)00841506103532	Threaded Peg, Non-Locking, 2.7mm x 30mm, Ti TPNL-27300-TS (01)00841506103617 (01)00841506103617
Threaded Peg, Non-Locking, 2.7mm x 16mm, Ti TPNL-27160-TS (01)00841506103549 (01)00841506103549	Threaded Peg, Non-Locking, 2.7mm x 32mm, Ti TPNL-27320-TS (01)00841506103624 (01)00841506103624
Threaded Peg, Non-Locking, 2.7mm x 18mm, Ti TPNL-27180-TS (01)00841506103556 (01)00841506103556	Threaded Peg, Non-Locking, 2.7mm x 34mm, Ti TPNL-27340-TS (01)00841506103907 (01)00841506103907
Threaded Peg, Non-Locking, 2.7mm x 20mm, Ti TPNL-27200-TS (01)00841506103563 (01)00841506103563	Threaded Peg, Non-Locking, 2.7mm x 36mm, Ti TPNL-27360-TS (01)00841506103914 (01)00841506103914
Threaded Peg, Non-Locking, 2.7mm x 22mm, Ti TPNL-27220-TS (01)00841506103570 (01)00841506103570	Threaded Peg, Non-Locking, 2.7mm x 38mm, Ti TPNL-27380-TS (01)00841506103921 (01)00841506103921
Threaded Peg, Non-Locking, 2.7mm x 24mm, Ti TPNL-27240-TS (01)00841506103587 (01)00841506103587	Threaded Peg, Non-Locking, 2.7mm x 40mm, Ti TPNL-27400-TS (01)00841506103938 (01)00841506103938

© 2026 Skeletal Dynamics, Inc

Wrist and Forearm Set

Inventory Control Sheet

Cortical Screw, Non-Locking	Cortical Screw, Locking
Screw, Cortical Non-Locking, 3.5mm x 8mm, Ti PANL-35080-TS (01)00841506102771 (01)00841506102771	Screw, Cortical, Locking, 3.5mm x 8mm, Ti COLS-35080-TS (01)00841506101071 (01)00841506101071
Screw, Cortical Non-Locking, 3.5mm x 9mm, Ti PANL-35090-TS (01)00841506102788 (01)00841506102788	Screw, Cortical, Locking, 3.5mm x 9mm, Ti COLS-35090-TS (01)00841506101088 (01)00841506101088
Screw, Cortical Non-Locking, 3.5mm x 10mm, Ti PANL-35100-TS (01)00841506102795 (01)00841506102795	Screw, Cortical, Locking, 3.5mm x 10mm, Ti COLS-35100-TS (01)00841506101095 (01)00841506101095
Screw, Cortical Non-Locking, 3.5mm x 11mm, Ti PANL-35110-TS (01)00841506102801 (01)00841506102801	Screw, Cortical, Locking, 3.5mm x 11mm, Ti COLS-35110-TS (01)00841506101101 (01)00841506101101
Screw, Cortical Non-Locking, 3.5mm x 12mm, Ti PANL-35120-TS (01)00841506102818 (01)00841506102818	Screw, Cortical, Locking, 3.5mm x 12mm, Ti COLS-35120-TS (01)00841506101118 (01)00841506101118
Screw, Cortical Non-Locking, 3.5mm x 13mm, Ti PANL-35130-TS (01)00841506102825 (01)00841506102825	Screw, Cortical, Locking, 3.5mm x 13mm, Ti COLS-35130-TS (01)00841506101125 (01)00841506101125
Screw, Cortical Non-Locking, 3.5mm x 14mm, Ti PANL-35140-TS (01)00841506102832 (01)00841506102832	Screw, Cortical, Locking, 3.5mm x 14mm, Ti COLS-35140-TS (01)00841506101132 (01)00841506101132
Screw, Cortical Non-Locking, 3.5mm x 15mm, Ti PANL-35150-TS (01)00841506102849 (01)00841506102849	Screw, Cortical, Locking, 3.5mm x 15mm, Ti COLS-35150-TS (01)00841506101149 (01)00841506101149
Screw, Cortical Non-Locking, 3.5mm x 16mm, Ti PANL-35160-TS (01)00841506102856 (01)00841506102856	Screw, Cortical, Locking, 3.5mm x 16mm, Ti COLS-35160-TS (01)00841506101156 (01)00841506101156
Screw, Cortical Non-Locking, 3.5mm x 18mm, Ti PANL-35180-TS (01)00841506102863 (01)00841506102863	Screw, Cortical, Locking, 3.5mm x 18mm, Ti COLS-35180-TS (01)00841506101163 (01)00841506101163
Screw, Cortical Non-Locking, 3.5mm x 20mm, Ti PANL-35200-TS (01)00841506104171 (01)00841506104171	Screw, Cortical, Locking, 3.5mm x 20mm, Ti COLS-35200-TS (01)00841506104034 (01)00841506104034
Screw, Cortical Non-Locking, 3.5mm x 22mm, Ti PANL-35220-TS (01)00841506104188 (01)00841506104188	Screw, Cortical, Locking, 3.5mm x 22mm, Ti COLS-35220-TS (01)00841506104041 (01)00841506104041
Screw, Cortical Non-Locking, 3.5mm x 24mm, Ti PANL-35240-TS (01)00841506104195 (01)00841506104195	Screw, Cortical, Locking, 3.5mm x 24mm, Ti COLS-35240-TS (01)00841506104058 (01)00841506104058

© 2026 Skeletal Dynamics, Inc

Wrist and Forearm Set
Inventory Control Sheet

Screw, Cortical Non-Locking, 3.5mm x 26mm, Ti PANL-35260-TS (01)00841506104201 (01) 00841506104201	Screw, Cortical, Locking, 3.5mm x 26mm, Ti COLS-35260-TS (01)00841506104065 (01) 00841506104065
Screw, Cortical Non-Locking, 3.5mm x 28mm, Ti PANL-35280-TS (01)00841506104218 (01) 00841506104218	Screw, Cortical, Locking, 3.5mm x 28mm, Ti COLS-35280-TS (01)00841506104072 (01) 00841506104072
Washer	
Washer, Button (Bronze) WBTN-HCLP (01)00841506105963 (01) 00841506105963	Washer, Button (Blue) WBTN-2750-T (01)00841506103730 (01) 00841506103730
General Single Use (Disposable) Instruments	
K-Wire, Standard Tip, 1.6mm x 127mm KWIR-STD-15127 (01)00841506102504 (01) 00841506102504	
General Reusable Instruments	
AIMing Guides, 1.5mm PDG-AIM-015 (01)00841506102870 (01) 00841506102870	
Hook Plate	
GEMINUS Hook Plate GMN-HP (01)00841506101514 (01) 00841506101514	GEMINUS Hook Plate, Screw GMN-HP-SCRW (01)00841506101545 (01) 00841506101545
Hook Plate – Polyaxial Screw, Locking (Cannulated)	
Screw, Polyaxial Locking, 2.5mm x 10mm Cannulated PALS-25100-CC (01)00841506102665 (01) 00841506102665	Screw, Polyaxial Locking, 2.5mm x 20mm Cannulated PALS-25200-CC (01)00841506102719 (01) 00841506102719
Screw, Polyaxial Locking, 2.5mm x 12mm Cannulated PALS-25120-CC (01)00841506102672 (01) 00841506102672	Screw, Polyaxial Locking, 2.5mm x 22mm Cannulated PALS-25220-CC (01)00841506102726 (01) 00841506102726
Screw, Polyaxial Locking, 2.5mm x 14mm Cannulated PALS-25140-CC (01)00841506102689 (01) 00841506102689	Screw, Polyaxial Locking, 2.5mm x 24mm Cannulated PALS-25240-CC (01)00841506102733 (01) 00841506102733
Screw, Polyaxial Locking, 2.5mm x 16mm Cannulated PALS-25160-CC (01)00841506102696 (01) 00841506102696	Screw, Polyaxial Locking, 2.5mm x 26mm Cannulated PALS-25260-CC (01)00841506102740 (01) 00841506102740
Screw, Polyaxial Locking, 2.5mm x 18mm Cannulated PALS-25180-CC (01)00841506102702 (01) 00841506102702	Screw, Polyaxial Locking, 2.5mm x 28mm Cannulated PALS-25280-CC (01)00841506102757 (01) 00841506102757
Hook Plate - Single Use (Disposable) Instruments	
K-Wire, Standard Tip, 0.9mm x 152mm KWIR-STD-09152 (01)00841506102498 (01) 00841506102498	PLS AIMing Guides PLS-AIM-0910 (01)00841506102887 (01) 00841506102887

IFU-02499-00 V.1 – February 2026
© 2026 Skeletal Dynamics, Inc

Wrist and Forearm Set

Inventory Control Sheet

Dorsal Spanning Plates	
Dorsal Spanning Plate, Short, Ti GMN-DSP-160 (01)00841506105307 (01)00841506105307	Dorsal Spanning Plate, Long, Ti GMN-DSP-210 (01)00841506101453 (01)00841506101453
Multi-Thread Screw, Locking	**Multi-Thread Screw, Compression**
Screw, Multi-Thread, Locking, 3.0mm x 6mm, Ti MTLS-30060-TS (01)00841506102511 (01)00841506102511	Screw, Multi-Thread, Compression, 3.0mm x 6mm, Ti MTNL-30060-TS (01)00841506102580 (01)00841506102580
Screw, Multi-Thread, Locking, 3.0mm x 8mm, Ti MTLS-30080-TS (01)00841506102528 (01)00841506102528	Screw, Multi-Thread, Compression, 3.0mm x 8mm, Ti MTNL-30080-TS (01)00841506102597 (01)00841506102597
Screw, Multi-Thread, Locking, 3.0mm x 10mm, Ti MTLS-30100-TS (01)00841506102535 (01)00841506102535	Screw, Multi-Thread, Compression, 3.0mm x 10mm, Ti MTNL-30100-TS (01)00841506102603 (01)00841506102603
Screw, Multi-Thread, Locking, 3.0mm x 12mm, Ti MTLS-30120-TS (01)00841506102542 (01)00841506102542	Screw, Multi-Thread, Compression, 3.0mm x 12mm, Ti MTNL-30120-TS (01)00841506102610 (01)00841506102610
Screw, Multi-Thread, Locking, 3.0mm x 14mm, Ti MTLS-30140-TS (01)00841506102559 (01)00841506102559	Screw, Multi-Thread, Compression, 3.0mm x 14mm, Ti MTNL-30140-TS (01)00841506102627 (01)00841506102627
Screw, Multi-Thread, Locking, 3.0mm x 16mm, Ti MTLS-30160-TS (01)00841506102566 (01)00841506102566	Screw, Multi-Thread, Compression, 3.0mm x 16mm, Ti MTNL-30160-TS (01)00841506102634 (01)00841506102634
Screw, Multi-Thread, Locking, 3.0mm x 18mm, Ti MTLS-30180-TS (01)00841506102573 (01)00841506102573	Screw, Multi-Thread, Compression, 3.0mm x 18mm, Ti MTNL-30180-TS (01)00841506102641 (01)00841506102641

IFU-02499-00 V.1 – February 2026
© 2026 Skeletal Dynamics, Inc